


Shoulder Pain – Acute

Post-surgery, acute injury (sports injury), medial shoulder instability, acute on chronic osteoarthritis, osteochondrosis (OC), osteochondrosis dissecans (OCD), supraspinatus tendinopathy, biceps brachii tendinopathy, biceps brachii tendon tear or avulsion fracture, post athletic activity.



1. A small or large Assisi Loop can be used to treat the shoulder joint in any size pet as long as it is positioned correctly. Positioning is extremely important for optimal effect.
2. Position the pet lying down, preferably flat on the ground or up on their elbows.
3. Identify the “point of the shoulder”.
4. Place the center of the Assisi Loop over the point of the shoulder. **See Pictures**
5. If using a LOOP-AID, attach the Assisi Loop with two Velcro straps to keep the Loop in place. **See Picture**
6. The signal generator can be positioned to help balance and stabilize the unit during treatment.
7. The signal generator should be located in view of the person treating the pet.
8. Do not leave the pet unattended when using this placement as small movements can change the positioning of the Loop and the area being treated.
9. Turn the unit on by pressing the button on the front of the white rubberized signal generator; green lights will illuminate about once per second.
10. Keep the unit in place for the full 15 minutes until the unit shuts off.
11. Use the unit 4 times daily, if possible, for 3-5 days. Treat twice daily for an additional 5 days or until pain is completely resolved.
12. There should be a minimum of 2 hours between treatments for optimal nitric oxide enhancement.
13. **For those treating osteoarthritis in multiple parts of the body:** The Loop is only guaranteed to last the *minimum* 150 treatments if it is not used more frequently than once every 2 hours: If you are treating OA in multiple parts of the body, you will need to treat EACH AREA at least twice per day. We recommend:
 - Using more than one Loop so you can treat multiple areas during the same 15 minute period
 - Use the same Loop for multiple areas, waiting 2 hours between treatments
 - Use the Loop more frequently than every 2 hours, knowing that the battery may last for *significantly fewer treatments and negate the warranty*



Effective Treatment Zone (ETZ): The 20 CM Assisi Loop has been designed to deliver a clinically effective therapeutic signal to damaged tissue that is positioned within the perimeter of a barrel-shaped treatment zone as illustrated.

Pet Owner Responsibility. A panel of licensed, experienced veterinarians has developed this Guide. Pet owners should use this protocol only to treat a condition listed above and as directed by a qualified veterinary clinician. Other protocols in this Guide should be used only as directed by a qualified veterinary clinician.

Important: Before starting treatment, please consult product label for optimal product use and safety guidelines. To learn more about the technology, visit our website.

Battery Life: When the lights are blinking 2-3 times per second, the battery is running low and a new Assisi Loop should be purchased. There should be a minimum of 2 hours between treatments for optimal battery efficiency.